

DOUBLE STAGE RO ULTRAPURE WATER SYSTEM WPS21-125UVF



DOUBLE STAGE RO ULTRAPURE WATER SYSTEM

WPS21-125UVF

System achieves water quality, little drain and low running cost. Applications like sample dilution, reagent preparation, microbiological analysis, water analysis and general HPLC makes this product an superior choice for water purification. Used in Laboratory, Manufacturing, Reefkeeping, Aquarium, Laboratory, Research.
Also known as Laboratory Double stage RO ultrapure Water Purification System.

WPS21-125UVF DOUBLE STAGE RO ULTRAPURE WATER SYSTEM



SPECIFICATIONS

Model	WPS21-125UVF
Feed Water Requirements*	
Water Inlet	Tap water: TDS<200 ppm (Extra pretreatment filter is recommended, if TDS>200 ppm)
Temperature	5-45°C
Pressure	1.0-4.0 Kg/cm ²
Flow Procedure**	PF+AC+RO+RO+(UV)+AC+DI+(UF)+TF
Bacteria	<0.1 cfu/ml
Output(25°C)****	1st stage RO water:125 L/hr, 2nd stage RO water: 60 L/hr
Pure water outlet	1st, 2nd stage RO and Ultrapure water
DimensionLxWxH	760x630x1190 mm
Weight	80 kg
Standard configuration	Main body (Including 1 set of cartridges) + built-in 2 tank (40L PE tank+2 gallon pressure tank)+ accessory bag
Power Consumption (W)	300 W
Power Supply	AC110-220 V, 50/60 Hz
Note	*The feed water quality will influence the pure water's quality and cartridges life-span. **PF:polypropylene spun fiber, AC:active carbon, RO:reverse osmosis, DI:ion exchange, UV:ultraviolet, UF:ultrafiltration, TF:terminal microfiltration. ***Value of number will be influenced by temperature and feed water quality. ****All the specifications are tested under the situation:feed water's TDS=200ppm, 25°C, 50psi and 15% recovery rate.
Ultrapure Water Quality	
Resistivity(25°C)	18.2 MΩ.cm
Heavy Metal Ion	<0.1 ppb

TOC***	<3 ppb
Particle (>0.2µm)	<1/ml
Endotoxin	<0.001 EU/ml
Rnases	<0.01 ng/ml
Dnases	<4pg/µl
1st stage RO water's TDS	TDS (ppm, mg/l) < TDS of tap water x 5%
2nd stage RO water's conductivity	1-5µs/cm, Organic rejection rate>99% (when MW>200 Dalton), Particles and bacteria rejection rate>99%

Centrifugen

Centrifugen

82 Wendell Avenue, STE 100, Pittsfield, MA, 01201, USA
Email: info@centrifugen.com | Website: centrifugen.com